PO BOX 100320 | COLUMBIA, SC 29202-3320 | DMEPDAC.COM | ISO 9001

PRICING DATA ANALYSIS AND CODING (PDAC)

PALMETTO GBA®
A CELERIAN GROUP COMPANY

February 10, 2024

KRISTEN TREASURE
MEDLINE INDUSTRIES LP
3 LAKES DR
NORTHFIELD, IL 60093

Document Control Number (DCN): 23353003000001

Manufacturer Name	Product Name	Model Number	Assigned HCPCS Code(s)
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD59U	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD59UH	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD59UZ	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD39U	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD39UH	A9900

Manufacturer Name	Product Name	Model Number	Assigned HCPCS Code(s)
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD39UZ	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD39U5C	A9900
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD69U	A4216
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD69UH	A4216
MEDLINE INDUSTRIES LP	ADDIPAK RESPIRATORY INHALATION SALINE SOLUTION	HUDRHUD69UZ	A4216

Dear KRISTEN TREASURE,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

A4216 STERILE WATER, SALINE AND/OR DEXTROSE, DILUENT/FLUSH, 10 ML

A9900 MISCELLANEOUS DME SUPPLY, ACCESSORY, AND/OR SERVICE COMPONENT OF ANOTHER HCPCS CODE

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at www.dmepdac.com. If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS). Further information for requesting updates to the PCL can be found on the PDAC website at www.dmepdac.com. It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions, please contact the PDAC HCPCS Helpline at (877) 735-1326 during the hours of 9:30 a.m. to 5:00 p.m. ET, Monday through Friday. You may also visit our website to chat with one of our representatives or select the Contact Us button at the top of the page for email, FAX or postal mail information.

Sincerely,

Pricing, Data Analysis, and Coding (PDAC)
Palmetto GBA, LLC
www.dmepdac.com